



• North
Atlantic
Area

[ARS Home](#) [About ARS](#) [Help](#) [Contact Us](#) [En Español](#)

 [Printable Version](#)  [E-mail this page](#)

Search

[Advanced Search](#)

Browse By Subject

▶ [Home](#)

▶ [About Us](#)

▶ [Research](#)

▶ [Products & Services](#)

◦ [Publications](#)

◦ [NAA Safety Health & Environmental](#)

◦ [SH&E](#)

◦ [NAA Employee Services](#)

◦ [EEO/Civil Rights](#)

▶ [People & Places](#)

▶ [News & Events](#)

▶ [Partnering](#)

▶ [Careers](#)

You are here: [Products & Services](#) / / Ventilation Equipment Quiz

Products & Services

Ventilation Equipments

Ventilation Equipment Quiz



[HOME](#)

First Name:

Last Name:

Student / Staff ID (email):

Supervisor:

The basic components of a biological safety cabinet are:

Stainless steel enclosure

HEPA filter

Motor Blowers

The physical barrier between product and worker is provided by:

Laminar Air Flow

Sash

Labcoat

HEPA Filter

It is permissible to use volatile chemicals in a biological safety cabinet.

True False

Baffles are:

Confused scientists

Movable panels on the back wall of a fumehood

Measurements of air flow drawn through the face of the hood to the hood exhaust

A biological safety cabinet should be disinfected with:

Ethanol

Bleach

Chemical which kills the organism

Turbulence of the laminar flow is caused by:

People walking by
Doors opening/closing
Clouds

To minimize turbulence, the researcher should:

Work uninterrupted
Put usage sign on door
Work alone

The proper sitting height @ the BSC places the armpit level at:

Work surface
Bottom of sash
Doesn't matter

Open flames are permitted in the BSC:

True False

Personal protective equipment required when using a BSC include:

NOTE! When finished, print out a copy, sign and date it, and return it to your Safety Officer.

SIGN YOUR NAME HERE:

Date:

Last Modified: 04/05/2007

[ARS Home](#) | [USDA.gov](#) | [Site Map](#) | [Policies and Links](#)

[FOIA](#) | [Accessibility Statement](#) | [Privacy Policy](#) | [Nondiscrimination Statement](#) | [Information Quality](#) | [USA.gov](#) | [White House](#)